


Mitutoyo

内側マイクロメータ用基準器 セラ内側マイクロチェッカ

CI-300C

CI-600C



ユーザーズマニュアル － 取扱説明書 －

ご使用前に本書をよくお読みのうえ、
正しくお使いください。お読みになった後は、
いつでも見られる所に必ず保管してください。

● 関連取扱説明書の読む順番は、「■ 本書の位置付け」
(2 ページ) を参照してください。

No. 99MBF609B
2025 年 6 月 1 日 発行 (1)



■ 本書の対象商品名および符号

商品名	符号
セラ内側マイクロチェッカ	CI-300C
	CI-600C

■ 本書に関するお願いとご注意

- 本書に記載の使用法によらない使用により損害が発生した場合は、弊社は一切その責任を負いかねます。
- 本商品を貸与または譲渡するときは、本書を本商品に添付してください。
- 本書を紛失または損傷されたときは、すみやかにお求めの販売店、または弊社営業・サービスの窓口（☎「営業・サービスの窓口」（App-1 ページ））にご相談ください。
- 本商品の操作の前に本書をよく読んでください。特に、「安全上のご注意」（5 ページ）、「取り扱い上のご注意」（5 ページ）の内容を十分に理解してから本商品をお使いください。
- 本書の内容は 2025 年 6 月現在の情報に基づいています。
- 本書の内容の一部または全部を転載・複製することは固くお断りいたします。
- 本書に掲載している図は、説明の都合上、強調や簡略化、または一部を省略していることがあります。また、機能理解と操作に支障を与えない範囲内で、実際の表示と異なることがあります。
- PDF 形式の取扱説明書を閲覧する場合は、Acrobat Reader を推奨します。
- 本文中の会社名、団体名、商品名などは、各社、各団体の商標、または登録商標です。

©2025 Mitutoyo Corporation. All rights reserved.

目次

目次	1
本書について	2
本書で使用されている表記	3
安全上のご注意	5
取り扱い上のご注意	5
輸出および非居住者への技術提供にあたってのご注意	6
欧州諸国などへの輸出に関するご注意	6
日本国内で本商品を廃棄する際のご注意	6
保証	7
免責	7
1 概要	9
1.1 特長.....	9
1.1.1 同梱物について	10
1.2 各部の名称.....	11
2 セットアップ	13
2.1 準備.....	14
2.2 補助ブロック・V 溝ボックスの取り付け.....	15
3 内側マイクロメータの基点確認	17
4 保守	19
5 仕様	23
5.1 仕様.....	23
5.2 外観寸法.....	24
5.3 標準付属品.....	25
5.4 オプション品.....	25
5.5 保証書.....	25
営業・サービスの窓口	App-1

本書について

■ 本書の位置付け

本書の位置付けと他の分冊との関係などを説明します。

セラ内側マイクロチェッカ
ユーザーズマニュアル（本書）

本商品の概要、使用方法や保守について説明しています。

内側マイクロメータのユーザーズマニュアル

本商品を使って基点確認を行う対象の測定器の取り扱いについて説明しています。下記符号の内側マイクロメータは対象外です。

IMJ-MJ、IMZ、IMZ-M、IMZ-MJ、IMS

ゲージブロック／セラブロック
ユーザーズマニュアル

ゲージブロック／セラブロックの使用方法、保守、定期校正／検査について説明しています。

■ 本書の対象読者

内側マイクロメータのご使用経験がある方を対象にしています。

また、測定器の検査方法を理解できることを前提とします。

本書で使用されている表記

■ 潜在的な危険性に対する注意喚起を示す表記

 危険	取り扱いを誤った場合に、「死亡または重傷を負う危険が切迫して生じることが想定される」内容を示します。
 注意	取り扱いを誤った場合に、「軽傷を負う可能性が想定される」内容を示します。
注記	取り扱いを誤った場合に、「物的損害の発生が想定される」内容を示します。
	引火注意 ガスに引火する危険性があることを示します。

■ 行為の禁止および行為の強制を示す表記

	行為の禁止の具体的な内容を示します。
	行為の強制の具体的な内容を示します。

■ 参考情報や参照先を示す表記

重要	本商品を使用するうえで知っておかなければならない情報を示します。
Tips	本文で説明している操作方法や手順に関連する詳細情報および参考情報を示します。
	本書または外部の取扱説明書に、参照すべき情報がある場合は、参照先を示します。 例：○○の詳細は、  「1.2 各部の名称」（11 ページ）を参照してください。

■ その他の表記

() (丸かっこ)	直前の内容の説明、補足説明を示します。
「 」 (かぎかっこ)	強調する語句を示します。また、参照文で参照先を示します。
1、2、3…	作業の順番と、その内容を示します。

■ 表記の使用例

重要情報を示します。

補足情報を示します。

安全情報を示します。

安全情報を示します。

番号とそれに続く文章が、実施する作業の順番と内容を示します。

先頭に「」マークがある一文は、操作結果を示します。

X 使用方法

X 使用方法

X.X 操作前の準備

X.X.X 電源ケーブルの接続

電源ケーブルの接続は、測定機本体の設置時に弊社またはお求めの販売店のサービス担当が行います。

重要
電源ケーブルは本商品に付属のものを使用してください。

Tips
非常時やメンテナンス時に、AC インレットから電源ケーブルを引き抜いてください。

X.X.X プローブの取り付け

プローブをプローブヘッドに取り付けます。ここではその接続手順を説明します。

注意
プローブの移動に伴い不慮の事故によるけがをするリスクがあります。

! プローブの取り付け作業は、測定機本体の電源を OFF にした状態で行ってください。

注記
入力信号が不安定となることで本商品に損傷を与えることがあります。

! プローブの取り付け作業は、入力信号を無効にした状態であっても、測定機本体の電源を OFF にした状態で行ってください。

1 プローブ背面のキー溝がアンロック位置にセットされていることを確認する
2 プローブヘッド下部のマークとプローブ上部のマークが合うようにセットする
3 専用の取り付け工具でプローブ背面のキー溝を回し、ロック位置にセットする

» プローブの取り付けが完了する。

XX
No. XXXXXXXXXX

X
使用
方法

安全上のご注意

ご使用の前に、この「安全上のご注意」をよくお読みのうえ、正しくお使いください。

ここに示した注意事項は、お使いになる人や、他の人への危害、財産への損害を未然に防ぐための内容を記載していますので、必ずお守りください。

危険



シンナーやベンジンなどの有機溶剤は、引火や自然発火による火災や、吸入による中毒などの危険性があります。



シンナーやベンジンなどの有機溶剤を使用しないでください。

注意

本商品やオプション品を移動する際、取り扱いを誤るとけがをするリスクがあります。



本商品が落ちたり、倒れたりすると、けがの原因になることがあります。慎重に取り扱いください。

取り扱い上のご注意

■ 本商品の用途や取り扱いについて

- 本商品は内側マイクロメータの基点確認のための機器です。
本商品を基点確認以外の目的で使用しないでください。
- 本商品は精密機器です。
取り扱いには十分注意してください。開梱、設置、使用、保管時に各部に衝撃を加えたり、無理な力をかけたりしないように十分注意してください。本商品に衝撃を与えた場合は、本商品の精度が損なわれるおそれがあります。

■ 使用環境

以下の場所で本商品をご使用ください。

- 塵やほこりの少ない場所
- 振動の少ない場所
- 推奨温度として 20 °C の環境下で温度変化が少ない場所
- 湿度の低い場所（70% 以下）

温度や湿度が急激に変化する以下のような場所での使用や保管はおやめください。

本商品の機能や測定結果に悪影響を与えるとともに、故障の原因になります。

- 切削油や水などの影響を直接受ける場所
- 日光、熱風、冷風（空調の風など）が直接当たる場所

■ 運搬・移動

本商品を輸送する際は、間隔ブロック（「1.2 各部の名称」（11 ページ））に防錆油を塗布し、防錆紙とビニールで包んでから出荷時のカートンで梱包してください。そのため、本梱包資材は大切に保管して再利用いただくか、輸送が必要になった場合、お求めの販売店、または弊社営業・サービスの窓口（「営業・サービスの窓口」（App-1 ページ））へご連絡ください。

■ 分解・改造

本商品は組み立ての際、精密に調整されています。

精度や機能に影響を与えるため、本書に記載されている内容以外の調整および分解・改造は絶対に行わないでください。保証の対象外となります。

■ お手入れについて

ご使用後は、清浄なエアーにて付着しているほこりなどを吹き飛ばして除去し、溶剤などを浸した繊維の出ないやわらかい布で汚れを拭き取り、間隔ブロックに液体防錆油を薄く塗布してください（防錆処理）。

シンナーやベンジンなどの有機溶剤は使用しないでください。

溶剤の推奨品：ノルマルヘプタン

液体防錆油の推奨品：モリコート スーパーグリス

詳細は、「4 保守」（19 ページ）を参照してください。

輸出および非居住者への技術提供にあたってのご注意

本商品は、「外国為替および外国貿易法の輸出貿易管理令別表第 1 もしくは外国為替令別表に定める 16 の項」によるキャッチオール規制貨物・キャッチオール規制技術（プログラムを含む）です。

本商品の輸出および日本国非居住者への技術提供にあたっては、経済産業省の許可が必要になる場合があります。

また、本商品に機能を追加するためにオプションの追加や改造を行った場合は、「外国為替および外国貿易法の輸出貿易管理令別表第 1 もしくは外国為替令別表に定める 1 から 15 の項」によるリスト規制貨物、リスト規制技術（プログラムを含む）に該当となることがあります。その場合の本商品の輸出および日本国非居住者への技術提供にあたっては、経済産業省の許可が必要になります。事前に弊社にご相談ください。

欧州諸国などへの輸出に関するご注意

本商品を輸出される際は、英文の取扱説明書・適合宣言書（場合によっては輸出国公用語）が必要となる場合があります。

詳細につきましては弊社にご相談ください。

日本国内で本商品を廃棄する際のご注意

- ・ 事業者として廃棄する場合は、廃棄物の処理および清掃に関する法律（廃棄物処理法）などの関連法令に従い、適正な廃棄処理をしてください。
- ・ 個人として廃棄する場合は、各自治体の廃棄ルールに従ってください。

保証

本商品は、厳重な品質管理のもとで製造されていますが、お客様の正常な使用状態において、万一お買い上げの日から1年以内に故障した場合は、無償で修理させていただきます。お求めの販売店、または弊社営業・サービスの窓口（☎「営業・サービスの窓口」（App-1 ページ））へご連絡ください。

次のような場合には、保証期間内でも有償修理となります。

- 使用による通常の損耗によって生じた故障および損傷
- メンテナンス上、修理上または取り扱い上の誤りおよび不当な改造による故障および損傷
- お買い上げ後の移動、落下あるいは輸送による故障および損傷
- 火災、塩害、ガス害、異常電圧、雷サージおよび天災地変などによる故障および損傷
- 高度に危険な活動に使用したことによる故障および損傷

本保証は日本国内において適切に設置され、本書に記載される指示に従って操作されている場合にのみ有効です。

本保証に規定される場合を除き、適用される法によって許される最大の範囲で、あらゆる性質の、すべての明示的・黙示的な条件、表明および保証（商品性に関する保証、特定の目的への適合性の保証、非侵害の保証または取引過程、使用または取引実務から生じる保証を含みますが、これらに限定されません）は、排除されます。

お客様は、お客様が意図された結果を実現するために本商品を選択したことによって生ずるすべての結果についての全責任を引き受けるものとします。

免責

ミットヨ、その関連会社およびそのサプライヤーは、いかなる場合においても、収益の損失、利益の損失、もしくはデータの損失、または本商品の使用もしくは使用不能によって生じた特別損害、直接損害、間接損害、派生的損害、付随的損害、または懲罰的損害について、原因および責任理論の如何にかかわらず、たとえミットヨ、その関連会社またはそのサプライヤーが当該損害の可能性について通知を受けていた場合であっても、責任を負いません。

前記にもかかわらず、ミットヨが、お客様による本商品の使用によって生じた損害または損失に対して責任があると判断された場合でも、いかなる場合においても、ミットヨ、その関連会社およびそのサプライヤーのお客様に対する責任は、契約に基づく、（過失を含む）不法行為とを問わず、本商品に対してお客様が支払った金額を超えないものとします。

国、州、または管轄地によっては、派生的損害または付随的損害に対する責任の排除または制限を認めていない場合があります。そのような国、州、または管轄地におけるミットヨの責任は、法に認められる最大の範囲内で排除または制限されるものとします。

1 概要


1

概要





1.1 特長

本商品は、内側マイクロメータ * の基点設定寸法の確認に用いる精密機器です。

* 使用可能な内側マイクロメータは、アンビル径が $\phi 8$ mm のもののみです（3 点式内側マイクロメータを除く）。




本商品の目盛は、本商品に補助ブロックを取り付けたときの補助ブロック間の寸法を示します。



1.1.1 同梱物について


セラ内側マイクロチェッカには、以下の構成部品が同梱されています。




番号	名称
①	セラ内側マイクロチェッカ本体
②	カラー
③	補助ブロック（セラミックス製）
④	V 溝ボックス
⑤	クランプねじまわし

重要

出荷時の梱包箱は、本商品を輸送する際の梱包材として再利用いただけます。大切に保管してください。特に梱包箱内のスペーサが紛失しないように注意してください。



1.2 各部の名称



番号	名称
①	基点ブロック（セラミックス製）
②	間隔ブロック（鋼製）
③	寸法設定ブロック（セラミックス製）

MEMO

2 セットアップ

2.1 準備	14
2.2 補助ブロック・V 溝ボックスの取り付け	15

2.1 準備

- 1 清浄なエアで本商品に付着しているほこりなどを吹き飛ばして除去する
- 2 基点ブロック、寸法測定ブロック、補助ブロックにかえりがある場合は、セラストン（オプション）で取り除く

重要

- ・セラストンに力を加えすぎると、寸法が狂う場合があります。打ち傷を作らないよう取り扱ってください。
- ・基点ブロックまたは寸法測定ブロックのかえりが除去できない場合は、修理が必要です。お求めの販売店、または弊社営業・サービスの窓口（☎「営業・サービスの窓口」（App-1 ページ））へご連絡ください。


- 3 本商品と内側マイクロメータの温度慣らしを行う

2.2 補助ブロック・V溝ボックスの取り付け

重要


V溝ボックス（2個）は側面に6桁の番号が刻印されています。必ず同じ番号のV溝ボックスを使用してください。他の番号のV溝ボックスと組み合わせて使用すると、測定誤差が生じる場合があります。

1 クランプねじが手前になるようにV溝ボックスを差し込む



2 補助ブロックを差し込む


隙間が狭い場合は、クランプねじを回して調整してください



3 V溝ボックスを指で矢印の方向に軽く押し付けて、クランプねじ（①）を軽く締める

重要

クランプねじ（①）は強く締めないでください。基点ブロックと寸法設定ブロックにひずみが発生するおそれがあります。




Tips

このとき、クランプねじ（②）はフリーの状態です。

4 左右それぞれのV溝ボックスを指で上から押し付けながら、クランプねじまわしでクランプねじを強く締め付ける

重要

- V溝ボックス下面と間隔ブロック上面の間に隙間ができると、正しい寸法を作ることができません。V溝ボックスは必ず指で上から押し付けてください。
- クランプねじは強く締め付けてください。締め付けが弱いと、測定中に補助ブロックが傾き、大きな測定誤差が生じます。
- クランプねじを締め付ける際は、クランプねじまわしが折れないように注意してください。




イラストは左側のV溝ボックスです。
右側のV溝ボックスも同様に実施してください。


3 内側マイクロメータの基点確認

測定範囲 125 mm ～ 150 mm の内側マイクロメータを例に説明します。


- 1** 内側マイクロメータ本体を基点設定寸法より 0.5 mm 程度小さくする




- 2** 調整アンビルとシンプル側のアンビルにカラーを取り付け、内側マイクロメータを V 溝ボックスの溝に載るように置く



- 3** 内側マイクロメータを矢印の方向に動かして、調整アンビルの測定面を補助ブロックに接触させる




- 4 シンブルを回転させ、シンブル側のアンビルの測定面が補助ブロックに接触したところで、内側マイクロメータの目盛を読み取る



- 5 本商品の目盛（125 mm）と、内側マイクロメータの読み取り値からゼロ点誤差を求める

- 6 内側マイクロメータを 0.5 mm 程度小さくして取り外し、左右の補助ブロックを取り外す



重要

ゼロ点誤差が大きい場合は、内側マイクロメータの調整が必要です。
調整方法の詳細は、 各内側マイクロメータの取扱説明書を参照してください。

4 保守

本商品の日常の点検、収納・保管について説明します。

本商品の性能を十分に発揮させ、安全に長期間で使用いただくために、以下の項目に従って保守を実施してください。

注意

本商品を移動する際、取り扱いを誤るとけがをするリスクがあります。



本商品が落ちたり、倒れたりすると、けがの原因になることがあります。慎重に取り扱ってください。

■ 日常の点検

汚れや部品不良などを確認してください。

汚れている場合は、清掃してください。部品不良の場合は、修理が必要です。お求めの販売店、または弊社営業・サービスの窓口（「営業・サービスの窓口」（App-1 ページ））へご連絡ください。

■ 清掃

ご使用後は、清浄なエアールにて付着しているほこりなどを吹き飛ばして除去し、無水アルコールなどの溶剤を浸した繊維の出ない布などで、油污などを拭き取ってください。

工場内設備のエアガンは、油分などが混入していることがありますので使用しないでください。

危険



シンナーやベンジンなどの有機溶剤は、引火や自然発火による火災や、吸入による中毒などの危険性があります。



シンナーやベンジンなどの有機溶剤を使用しないでください。

注記

洗剤、シンナーやベンジンなどの有機溶剤は品質の劣化や損傷を与えますので、使用しないでください。

■ 収納・保管

清掃後に間隔ブロックに防錆油を塗布し、防錆紙とビニールで包んでから出荷時のカートンに保管してください。（長期保管の場合も同様）

重要

保管時は、温度・湿度変化の激しい場所を避けて、恒温恒湿な状態で保管してください。

Tips

基点ブロックと寸法設定ブロックはセラミックス製のため、防錆処理は不要です。

■ 校正・修理

校正および修理依頼時は、出荷時のカートンや緩衝材で梱包し、お求めの販売店、または弊社営業・サービスの窓口（「営業・サービスの窓口」（App-1 ページ））へご連絡ください。

■ 再梱包

以下の手順で再梱包してください。


- 1 本商品を防錆紙とビニールで包み、布輪を掛ける



- 2 布輪を持って、梱包箱に収める



- 3 梱包箱内のスペーサが外れていないことを確認する



- 4 押さえ（緩衝材）をセットする



5 上板をセットする**6** 梱包箱を閉じる

MEMO

5 仕様

5.1 仕様	23
5.2 外観寸法	24
5.3 標準付属品	25
5.4 オプション品	25
5.5 保証書	25

5.1 仕様

コード No.	515-585	515-586
符号	CI-300C	CI-600C
測定範囲	25 mm ～ 300 mm	25 mm ～ 600 mm
ブロックの寸法許容差	$\pm \left(1 + \frac{L}{150}\right) \mu\text{m}$ L: ブロック間の長さ (mm)	
質量	3.9 kg	7.0 kg
使用環境	温度：5 °C ～ 40 °C	
保存環境	温度：-10 °C ～ 60 °C	


重要

底面と測定面の直角度は出ていません。

5.2 外觀寸法


● CI-300C

単位：mm



● CI-600C

単位：mm



5.3 標準付属品

パーツ No.	品名	数量	備考
613671-022	補助ブロック	2	セラブロック (10 mm)
940286	V 溝ボックス	1 対 (2 個)	
602195	カラー	2	ø8 用
600324	クランプねじまわし	1	
99MBF609B	ユーザズマニュアル (本書)	1	
—	検査成績書	1	

5.4 オプション品

パーツ No. / コード No.	品名
602160	格納木箱 (CI-300C 用)
602163	格納木箱 (CI-600C 用)

5.5 保証書

Mitutoyo		WA100 03
保 証 書		
商品名	コード番号	
製造番号	お買い上げ日	
〈お願い〉 お買い上げの際、上記各欄の必要な項目についてご記入ください。 当社は、品質保証の国際規格「ISO9000シリーズ」に準づく品質システムにより品質保証をおこなっており、この商品は、ミツトヨ検査規格に適合し、使用した標準器は国家標準へのトレーサビリティを確立しております。 取扱説明書、本体貼付ラベルなどの注意書に従った正常な使用状態で保証期間内に故障した場合には、本書記載内容により無料修理をさせていただきます。		
保証期間 (お買い上げ日より) 本 体 : 1年		株式会社 ミツトヨ 神奈川県横浜市高津区坂戸1-20-1 〒213-8501 電話 (044) 813-8201 (代表)
■ 保証期間内でも次のような場合には有料修理となります。 ◆ 使用上の誤り、改造や不当な修理による故障または損傷 ◆ お買い上げ後の移設、輸送、落下などによる故障または損傷 ◆ 不適当な保守、保管、保存による故障または損傷 ◆ 異常電圧、指定外の使用電源 (電圧、周波数) による故障または損傷 ◆ 消耗品の消耗による故障または損傷 ◆ 火災、地震、水害、落雷、その他の天災地変、公害、煙害、ガス害 (硫化ガスなど) による故障または損傷 ◆ 本書のご提示がない場合 ◆ その他当社の責任とみなされない故障または損傷 ■ 修理は、お買い上げの販売店または最寄りの弊社営業所へご依頼ください。 ※ 本保証書はお買い上げいただいた国においてのみ有効です。 ※ 本保証書は再発行いたしませんので大切に保管してください。		

営業・サービスの窓口

2025 年 1 月現在

千歳事務所	千歳市東雲町 2-28 プレミアテラス東雲 1F 電話：(0123) 25-5886	〒 066-0042 ファクス：(0123) 25-5887
盛岡事務所	盛岡市前九年 1-4-56 グローリープレーステナント 102 電話：(019) 601-9350	〒 020-0127 ファクス：(019) 601-9351
仙台営業所	仙台市若林区卸町東 1-7-30 電話：(022) 231-6881	〒 984-0002 ファクス：(022) 231-6884
郡山営業所	郡山市朝日 2-1-5 丸久ビル 103 号 電話：(024) 931-4331	〒 963-8024 ファクス：(024) 931-4332
宇都宮営業所	宇都宮市平松本町 796-1 電話：(028) 660-6240	〒 321-0932 ファクス：(028) 660-6248
水戸営業所	水戸市元吉田町 260-3 電話：(029) 303-5371	〒 310-0836 ファクス：(029) 303-5372
新潟営業所	新潟市中央区新和 1-6-10 リファール新和 1 階 B 号室 電話：(025) 281-4360	〒 950-0972 ファクス：(025) 281-4367
伊勢崎営業所	伊勢崎市東小保方町 3519 雄宝ビル 102 電話：(0270) 75-3756	〒 379-2234 ファクス：(0270) 75-3768
さいたま営業所	さいたま市北区宮原町 3-429-1 電話：(048) 667-1431	〒 331-0812 ファクス：(048) 667-1434
川崎営業所	川崎市高津区坂戸 1-20-1 電話：(044) 813-1611	〒 213-8533 ファクス：(044) 813-1610
東京営業所	港区芝 4-3-14 仏教伝道センタービル 5F 電話：(03) 3452-0481	〒 108-0014 ファクス（川崎営業所）：(044) 813-1610
厚木営業所	厚木市中町 2-6-10 東武太朋ビル 2F 電話：(046) 259-6400	〒 243-0018 ファクス：(046) 259-6404
富士事務所	富士市伝法 1647 電話（厚木営業所）：(046) 259-6400	〒 417-0061 ファクス（厚木営業所）：(046) 259-6404
諏訪営業所	諏訪市中洲 582-2 電話：(0266) 53-6414	〒 392-0015 ファクス：(0266) 58-1830
上田事務所	上田市材木町 2-9-4 YCC ビル 1F 電話（諏訪営業所）：(0266) 53-6414	〒 386-0014 ファクス（諏訪営業所）：(0266) 58-1830
浜松営業所	浜松市中央区和田町 587-1 電話：(053) 464-1451	〒 435-0016 ファクス：(053) 464-1683
安城営業所	安城市住吉町 5-19-5 電話：(0566) 98-7070	〒 446-0072 ファクス：(0566) 98-6761
中部オートモーティブ営業所	安城市住吉町 5-19-5 電話：(0566) 98-7070	〒 446-0072 ファクス：(0566) 98-6761

名古屋営業所	名古屋市昭和区鶴舞 4-14-26 電話：(052) 741-0382	〒 466-0064 ファクス：(052) 733-0921
金沢営業所	金沢市駅西本町 6-6-13 電話：(076) 222-1160	〒 920-0025 ファクス：(076) 222-1161
大阪営業所	大阪市住之江区南港北 1-4-34 電話：(06) 6613-8801	〒 559-0034 ファクス：(06) 6613-8817
神戸営業所	神戸市西区丸塚 1-25-15 電話：(078) 924-4560	〒 651-2143 ファクス：(078) 924-4562
京滋営業所	草津市大路 2-13-27 辻第 3 ビル 1F 電話：(077) 569-4171	〒 525-0032 ファクス：(077) 569-4172
岡山営業所	岡山市北区田中 134-107 電話：(086) 242-5625	〒 700-0951 ファクス：(086) 242-5653
広島営業所	東広島市八本松東 2-15-20 電話：(082) 427-1161	〒 739-0142 ファクス：(082) 427-1163
高知事務所	香南市香我美町徳王子 120-1 電話：(0887) 52-9470	〒 781-5332 ファクス：(0887) 52-9490
福岡営業所	福岡市博多区博多駅南 4-16-37 電話：(092) 411-2911	〒 812-0016 ファクス：(092) 473-1470
熊本事務所	熊本市北区弓削 2-12-33 電話（福岡営業所）：(092) 411-2911	〒 861-8002 ファクス（福岡営業所）：(092) 473-1470
センシング営業課	川崎市高津区坂戸 1-20-1 電話：(044) 813-8236	〒 213-8533 ファクス：(044) 822-8140
地震機器課	川崎市高津区坂戸 1-20-1 電話：(044) 455-5021	〒 213-8533 ファクス：(044) 822-8140

◆ 商品の故障および操作方法に関してのご相談・お問い合わせ

カスタマーサポートセンタ 電話：(0570) 073214 ファクス：(044) 813-1691

- 受付時間：弊社営業日 8:30 ～ 12:00 13:00 ～ 17:15（16:45*） * 月末最終営業日
- お願い： 測定機器商品（形状測定機、硬さ試験機、座標計測機器、画像測定機、光学機器ほか）に関するお問い合わせの場合は、該当商品のコード番号、シリアル番号をお手元にご用意ください。

改訂履歴

発行年月日	版数	改訂内容
2025 年 6 月 1 日	初版	発行

株式会社 ミットヨ

神奈川県川崎市高津区坂戸 1-20-1 〒 213-8533


ホームページ : <https://www.mitutoyo.co.jp>



Reference Instrument for Inside Micrometers CERA Inside Micro Checker

CI-300C

CI-600C



User's Manual - Instructions for use -

Read this document thoroughly before
operating the product. After reading, retain it
close at hand for future reference.
This English language version of the document
contains the original instructions.

- For the reading order of the related User's Manuals,
see "■ Positioning of this document" on page 2.

No. 99MBF609B
Date of publication: June 1, 2025 (1)



■ Product names and model numbers covered in this document

Product name	Model number
CERA Inside Micro Checker	CI-300C CI-600C

■ Notice regarding this document

- Mitutoyo Corporation assumes no responsibilities for any damage to the product, caused by its use not conforming to the procedure described in this document.
- Upon loan or transfer of this product, be sure to attach this document to the product.
- In the event of loss or damage to this document, immediately contact the agent where you purchased the product or a Mitutoyo sales office (📖 "SERVICE NETWORK" on page App-1).
- Read this document thoroughly before operating the product. In particular, be sure to fully understand "Safety Precautions" on page 5 and "Precautions for Use" on page 5.
- The contents of this document are based on information current as of June 2025.
- No part or whole of this document may be transmitted or reproduced by any means without prior written permission of Mitutoyo Corporation.
- Some figures in this document may be highlighted, simplified or partially omitted for convenience of explanation. In addition, some of them may differ from actual ones to the extent that no user will misunderstand the functions and operations.
- Acrobat Reader is recommended for viewing PDF user's manuals.
- The corporation, organization and product names that appear in this document are their trademarks or registered trademarks.

©2025 Mitutoyo Corporation. All rights reserved.

Contents

Contents	1
About This Document	2
Conventions Used in This Document	3
Safety Precautions	5
Precautions for Use	5
Export Control Compliance	6
Notes on Export to European Countries	6
Disposal of Products outside the European Countries	6
Warranty	7
Disclaimer	7
1 Overview	9
1.1 Features	9
1.2 Name of Each Part	11
2 Setup	13
2.1 Preparation	14
2.2 Attaching the Auxiliary Blocks and the Support Clamp	15
3 Checking the Zero-Point of the Inside Micrometer	17
4 Maintenance	19
5 Specifications	23
5.1 Specifications	23
5.2 External Dimensions	24
5.3 Standard Accessories	25
5.4 Optional Accessories	25
5.5 Warranty Card	25
SERVICE NETWORK	App-1

About This Document

■ Positioning of this document

This describes the positioning of this document and its relationship with other installments.

CERA Inside Micro Checker
User's Manual (This document)

Provides an overview of this product as well as a description of how to use the product and perform maintenance.

Inside Micrometer User's Manual

Provides a description of how to operate the measuring instrument on which this product is to be used to perform a zero-point check.
Inside micrometers of the following series are not subject to zero-point checking.
IMJ-MJ, IMZ, IMZ-M, IMZ-MJ, and IMS

Gauge Block / CERA Block
User's Manual

Provides a description of how to use gauge blocks and CERA blocks as well as of the execution of maintenance, periodic calibrations, and inspections.

■ Intended readers of this document

This document is intended for operators who have already operated an inside-micrometer.
They are also assumed to be able to understand inspection procedure of measuring instruments.

Conventions Used in This Document

■ Safety reminder conventions warning against potential hazards

 DANGER	Indicates a hazard with a high level of risk which, if not avoided, will result in death or serious injury .
 CAUTION	Indicates a hazard with a low level of risk which, if not avoided, could result in minor or moderate injury .
NOTICE	Indicates a situation which, if not avoided, may result in property damage .
	Flammable material Alerts the user to a specific hazardous situation that means "Caution, risk of igniting gas".

■ Conventions indicating prohibited and mandatory actions

	Indicates concrete information about prohibited actions.
	Indicates concrete information about mandatory actions.

■ Conventions indicating referential information or reference location

IMPORTANT	Indicates information that must be known when using the product.
Tips	Indicates further information and details relevant for the operating methods and procedures that are explained in that section.
	Indicates reference location if there is information that should be referred to in this document or an extraneous User's Manual. Example: For details about XX, see  "1.2 Name of Each Part" on page 11.

Other conventions

(): Round brackets	Represent a paraphrase of an immediately preceding phrase or a supplementary explanation.
" ": Double quotation marks	Represent a highlighted phrase. They also indicate an index where information to be referenced is described.
1, 2, 3...	Indicates the order and the contents of tasks.

Example of conventions use

Important information

Supplementary information

Indicates an alert.

Indicates an alert.

Indicates an operating procedure to be performed or its outline.

A "»" mark at the beginning of a sentence indicates the result of an operation.

X Operation

X Operation

X.X Precautions for Use

X.X.X Connecting the Power Cable

Mitutoyo service personnel or the agent where you purchased the product will connect the power cable at the installation of the main unit.

IMPORTANT

Use only the power cable that is supplied with this product.

Tips

In case of emergency or for maintenance purposes, after unplugging the power cable from the AC inlet, plug it in again.

X.X.X Mounting the Probe

Mount the probe onto the probe head by "auto joint connection". This section explains connection procedure.

CAUTION

There is a risk of injury due to accidents that accompany the movements of the probe.

! Turn off the main unit before mounting the probe.

NOTICE

There is a risk of damage to this product due to the instability of input signals.

! Turn off the main unit before mounting the probe even if the input signals are disabled.

1 Make sure that the keyway on the rear of the probe is set to the unlock position.

2 Align the marks on the probe head and the probe.

3 Using the accessory tool, turn the keyway on the rear of the probe to the lock position.

» The probe mounting is completed.

XX

No. XXXXXXXXXX

X
Operation

Safety Precautions

Read these "Safety Precautions" thoroughly before operating the product to use it properly. These safety precautions include such information as to prevent injury to the operator and other persons, damage to property and product defects. Be sure to observe these precautions carefully.

DANGER



Organic solvents such as diluents and benzine spirits may catch fire either due to being ignited or due to combusting spontaneously. Inhaling them also poses a risk of poisoning.



Do not use organic solvents such as diluents or benzine spirits.

CAUTION

Handling this product or its optional accessories improperly during transport poses a risk of injuries.



Falling or toppling of the product may lead to injuries. Handle with care.

Precautions for Use

■ Use and handling of the product

- This product is an instrument that is used for checking the zero-point of inside micrometers. Do not use this product for purposes other than for zero-point checking.
- The product is a precision instrument.
Handle with sufficient care. During the unpacking, installation, operation, and storage of this product, take sufficient care that its parts do not sustain any shocks and that no excessive force is applied. If this product sustains any shocks, its measuring accuracy may become impaired.

■ Environment for operation

Use the product in the following places.

- Where there is minimal dust and grit
- Where there are minimal vibrations
- An environment that varies minimally from the recommended temperature of 20 °C
- An environment in which the humidity is low (70% or less)

Do not use or store the product in the following places where the temperature and humidity drastically fluctuate.

The product's functions and measurement results will be adversely affected and it will cause malfunction.

- An environment that directly receives the influence of cutting fluid, water, etc.
- An environment that directly receives sunlight as well as both hot and cold airflows (such as those from an air conditioner)

■ Transport

Before transporting this product, apply corrosion protective lubricant to the spacing blocks (📖 "1.2 Name of Each Part" on page 11), and wrap a piece of corrosion protective packaging paper and a plastic packaging sheet around the product. After that, the product is to be packed into the cardboard box that is used for shipping. For that reason, the packing materials are to be securely stored and re-used. However, please contact the agent where you purchased the product or a Mitutoyo sales office (📖 "SERVICE NETWORK" on page App-1) if you require the delivery of additional packing materials.

■ Disassembly and modification

This product is adjusted precisely during its assembly.

Performing disassembly, modification, or any adjustments other than those described in this document will affect the measuring accuracy and functions of this product. Therefore, please refrain from doing so. Doing so will void the warranty.

■ About daily care

After each use, remove any adhering particles such as dust by blowing them away with clean air, and then wipe off any grime by using a soft, lint-free cloth that has been moistened with a solvent. After that, apply liquid-type corrosion protective lubricant in a thin layer to the spacing blocks (to prevent rusting). Do not use organic solvents such as diluents or benzine spirits.

Recommended solvent: Normal heptane

Recommended liquid-type corrosion protective lubricant: MOLYKOTE® Supergliss

For details, see 📖 "4 Maintenance" on page 19.

Export Control Compliance

This product falls into the Catch-All-Controlled Goods and/or Catch-All-Controlled Technologies (including Programs) under Category 16 of Appended Table 1 of Export Trade Control Order or under Category 16 of Appended Table of Foreign Exchange Control Order, based on Foreign Exchange and Foreign Trade Act of Japan.

If you intend re-export of the product from a country other than Japan, re-sale of the product in a country other than Japan, or re-providing of the technology (including Programs), you shall observe the regulations of your country.

Also, if an option is added or modified to add a function to this product, this product may fall under the category of List-Control Goods, List-Control Technology (including Programs) under Category 1 - 15 of Appended Table 1 of Export Trade Control Order or under Category 1 - 15 of Appended Table of Foreign Exchange Control Order, based on Foreign Exchange and Foreign Trade Act of Japan. In that case, if you intend re-export of the product from a country other than Japan, re-sale of the product in a country other than Japan, or re-providing of the technology (including Programs), you shall observe the regulations of your country. Please contact Mitutoyo in advance.

Notes on Export to European Countries

When you intend exporting of this product to any of the European countries, it may be required to provide User's Manual(s) in English and Declaration of Conformity in English (in some cases, the official language of the country to be exported).

For detailed information, please contact Mitutoyo in advance.

Disposal of Products outside the European Countries

Please follow the official instruction in each community and country.

Warranty

This product has been manufactured under strict quality management, but should it develop problems within one year of the date of purchase in normal use, repair shall be performed free of charge. Please contact the agent where you purchased the product or Mitutoyo sales representative ( "SERVICE NETWORK" on page App-1).

If this product fails or is damaged for any of the following reasons, it will be subject to a repair charge, even if it is still under warranty.

- Failure or damage owing to fair wear and tear
- Failure or damage owing to inappropriate handling, maintenance or repair, or to unauthorized modification
- Failure or damage owing to transport, dropping, or relocation of the product after purchase
- Failure or damage owing to fire, salt, gas, abnormal voltage, lightning surge, or natural disaster
- Failure or damage owing to use in ultra-hazardous activities

This warranty is effective only where the product is properly installed and operated in conformance with the instructions in this document within the original country of the installation.

EXCEPT AS SPECIFIED IN THIS WARRANTY, ALL EXPRESS OR IMPLIED CONDITIONS, REPRESENTATIONS, AND WARRANTIES OF ANY NATURE WHATSOEVER INCLUDING, WITHOUT LIMITATION, ANY IMPLIED WARRANTY OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, NONINFRINGEMENT OR WARRANTY ARISING FROM A COURSE OF DEALING, USAGE, OR TRADE PRACTICE, ARE HEREBY EXCLUDED TO THE MAXIMUM EXTENT ALLOWED BY APPLICABLE LAW.

You assume responsibility for all results due to the selection of this product to achieve your intended results.

Disclaimer

IN NO EVENT WILL MITUTOYO, ITS AFFILIATED AND RELATED COMPANIES AND SUPPLIERS BE LIABLE FOR ANY LOST REVENUE, PROFIT, OR DATA, OR FOR SPECIAL, DIRECT, INDIRECT, CONSEQUENTIAL, INCIDENTAL, OR PUNITIVE DAMAGES HOWEVER CAUSED AND REGARDLESS OF THE THEORY OF LIABILITY ARISING OUT OF THE USE OF OR INABILITY TO USE THIS PRODUCT EVEN IF MITUTOYO OR ITS AFFILIATED AND RELATED COMPANIES AND/OR SUPPLIERS HAVE BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES.

If, notwithstanding the foregoing, Mitutoyo is found to be liable to you for any damage or loss which arises out of or is in any way connected with use of this product by you, in no event shall Mitutoyo's and/or its affiliated and related companies' and suppliers' liability to you, whether in contract, tort (including negligence), or otherwise, exceed the price paid by you for the product only.


The foregoing limitations shall apply even if the above-stated warranty fails of its essential purpose. BECAUSE SOME COUNTRIES, STATES OR JURISDICTIONS DO NOT ALLOW THE EXCLUSION OR THE LIMITATION OF LIABILITY FOR CONSEQUENTIAL OR INCIDENTAL DAMAGES, IN SUCH COUNTRIES, STATES OR JURISDICTIONS, MITUTOYO'S LIABILITY SHALL BE LIMITED TO THE EXTENT PERMITTED BY LAW.

1 Overview





1.1 Features

This product is a precision instrument that is used to check the zero-point of inside micrometers*.

* Only inside micrometers that have an anvil diameter of $\varnothing 8$ mm are testable (excluding three-point inside micrometers).




The graduations marked on this product indicate the distance between the auxiliary blocks that are attached to the product.



1.1.1 Included Components


The following components are included with the CERA Inside Micro Checker.




No.	Name
①	CERA inside micrometer checker main unit
②	Collar
③	Auxiliary block (ceramic)
④	Support clamp
⑤	Clamp screwdriver

IMPORTANT

The original packaging box can be reused as packaging material for transporting this product. Please store it in a safe place.
Take special care not to lose the spacers that are in the cardboard packing box.



1.2 Name of Each Part



No.	Name
①	Zero-point block (ceramic)
②	Spacing block (steel)
③	Dimension-setting block (ceramic)

MEMO

2 Setup

2.1 Preparation 14

2.2 Attaching the Auxiliary Blocks and the Support Clamp 15

2.1 Preparation

- 1 Remove any particles such as dust adhering to this product by blowing them away with clean air.
- 2 If burrs have been found on the zero-point block, a dimension-measuring block, or an auxiliary block, remove them by using a Ceraston (optional accessory).

IMPORTANT

- Applying too much force to the Ceraston may affect the dimensions of the block. Be careful not to dent or scratch the surfaces of the block.
- In cases where it is not possible to eliminate the burrs found on the zero-point block or dimension-measuring block, a repair is required. Please contact the agent where you purchased the product or a Mitutoyo sales office (☎ "SERVICE NETWORK" on page App-1).


- 3 Allow enough time for this product and the inside micrometer to acclimate to room temperature.

2.2 Attaching the Auxiliary Blocks and the Support Clamp

IMPORTANT


A six-digit number is engraved on one side of each support clamp (two pieces). Always use a pair of support clamps that are engraved with the same number. Using support clamps that are engraved with different numbers may lead to measurement errors.

- 1 Insert the support clamps so that their clamp screws face toward you.



- 2 Insert the auxiliary blocks.


If the insertion spaces are too narrow, adjust them by rotating the clamp screws.



- 3 While using your fingers to gently push the support clamps in the direction indicated by the arrows, lightly tighten the clamp screws (①).

IMPORTANT

Do not overtighten the clamp screws (①). Otherwise, the zero-point block or dimension-setting block may become deformed.




Tips

At this stage, the clamp screws (②) are left loose.

- 4** While using your fingers to hold down the left-side support clamp and right-side support clamp one at a time, firmly tighten the clamp screw with the clamp screwdriver.

IMPORTANT

- If there is a gap between the bottom surface of each support clamp and the top surface of the spacing block, it will not be possible to obtain the correct dimensions. Be sure to use your fingers to hold down the support clamp.
- Firmly tighten the clamp screws. If the screws have not been tightened firmly enough, the auxiliary blocks will tilt during measurement and thereby cause significant measurement errors.
- While tightening the clamp screws, be careful not to break the clamp screwdriver.




This figure shows the left-side support clamp.
Perform the same work on the right-side support clamp.


3 Checking the Zero-Point of the Inside Micrometer

In this chapter, an inside micrometer with a measurement range of 125 mm to 150 mm is used as an example.


- 1 Adjust the inside micrometer main unit so that its length becomes approximately 0.5 mm shorter than the zero-point setting dimension.




- 2 Attach one collar piece to the adjustment anvil and to the anvil on the thimble side, and then place the inside micrometer on the support clamps.



- 3 Bring the measurement surface of the adjustment anvil into contact with the auxiliary block by moving the inside micrometer in the direction indicated by the arrow.




- 4 Rotate the thimble until the measurement surface of the anvil on the thimble side comes into contact with the auxiliary block, and then check the reading on the inside micrometer.



- 5 Calculate the difference (zero-point error) between the graduation mark on this product (125 mm) and the reading on the inside micrometer.

- 6 Decrease the length of the inside micrometer by approximately 0.5 mm and remove it, then remove the left and right auxiliary blocks.



IMPORTANT

If the zero-point error is large, the inside micrometer is in need of adjustment.

For details about how to make adjustments, see  the user's manual for each inside micrometer.

4 Maintenance

This chapter describes the daily inspection and storage of this product.

In order to obtain optimum performance and to ensure safe use of this product over an extended period of time, please observe the following guidelines when performing maintenance.

CAUTION

Handling this product improperly during transport poses a risk of injuries.



Falling or toppling of the product may lead to injuries. Handle with care.

■ Daily inspection

Check for the presence of any grime or parts defects.

If the product is dirty, clean it. If any parts defects are found, they must be repaired. Please contact the agent where you purchased the product or a Mitutoyo sales office ( "SERVICE NETWORK" on page App-1).

■ Cleaning

After each use, remove any adhering particles such as dust by blowing them away with clean air, and then wipe off any grime such as oil stains by using a lint-free cloth that has been moistened with a solvent such as absolute alcohol.

Due to the risk that the air-spray guns furnished in the factory may been contaminated with oil or the like, they are not to be used.

DANGER



Organic solvents such as diluents and benzine spirits may catch fire either due to being ignited or due to combusting spontaneously. Inhaling them also poses a risk of poisoning.



Do not use organic solvents such as diluents or benzine spirits.

NOTICE

Do not use detergents or organic solvents such as diluents or benzine spirits because they will adversely affect product quality and cause damage to the product.

■ Storage

After cleaning, apply corrosion protective lubricant to the spacing blocks, and then wrap a piece of anti-corrosion packaging paper and a plastic packaging sheet around the product. After that, the product is to be stored in the cardboard box that is used for shipping. (This is the same as the procedure for long-term storage.)

IMPORTANT

Store the product in an environment with a constant temperature and humidity, and avoid any environments where there are drastic changes in temperature or humidity.

Tips

The zero-point block and dimension-setting blocks are ceramic parts, so they do not require anti-rust treatment.

■ Calibration and repair

When consigning this product for calibration or repair, pack it in the cardboard box that is used for shipping along with cushioning materials. After that, please contact the agent where you purchased the product or a Mitutoyo sales office (☞ "SERVICE NETWORK" on page App-1).

■ Repacking

When repacking this product, follow the procedure below.


- 1** Wrap a piece of anti-corrosive packaging paper and plastic packaging sheet around the product and then tie a loop of fabric onto it.



- 2** Hold the packaged product by the loop of fabric and pack it into the cardboard packing box.



- 3** Check that the spacer has not deviated from its position in the cardboard packing box.



4 Set the cushioning materials.**5 Set the top inner cover.****6 Seal the cardboard packing box.**

MEMO

5 Specifications

5.1 Specifications	23
5.2 External Dimensions	24
5.3 Standard Accessories	25
5.4 Optional Accessories	25
5.5 Warranty Card	25

5.1 Specifications

Code No.	515-585	515-586
Model number	CI-300C	CI-600C
Measurement range	25 mm to 300 mm	25 mm to 600 mm
Block pitch accuracy	$\pm \left(1 + \frac{L}{150} \right) \mu\text{m}$ L: Length to check (mm)	
Mass	3.9 kg	7.0 kg
Environment for operation	Temperature: 5 °C to 40 °C	
Environment for storage	Temperature: -10 °C to 60 °C	


IMPORTANT

The perpendicularity between the bottom surface and the measurement surface is not specified.

5.2 External Dimensions


● CI-300C

Unit: mm



● CI-600C

Unit: mm




5.3 Standard Accessories

Part No.	Part name	Qty.	Remarks
613671-022	Auxiliary block	2	CERA block (10 mm)
940286	Support clamp	One pair (2 pcs.)	
602195	Collar	2	For ø8
600324	Clamp screwdriver	1	
99MBF609B	User's Manual (this document)	1	
-	Certificate of inspection	1	

5.4 Optional Accessories

Part No. / Code No.	Part name
602160	Wooden box (for CI-300C)
602163	Wooden box (for CI-600C)

5.5 Warranty Card

		Mitutoyo Corporation 20-1, Sakado 1-chome, Takatsu-ku, Kawasaki-shi, Kanagawa 213-8533, Japan Phone (044) 813-8230	
ONE YEAR WARRANTY			
Product Name		Code No.	
Serial No.		Date Purchased	
Mitutoyo Corporation conducts an own Quality management system based on the requirements of ISO 9001. This product conforms to the Mitutoyo inspection standard, and the standard (s) used for the calibration is /are traceable to the national standard (s). In the event that this product should fail within one (1) year from the original date of purchase through normal use conforming to the User's Manual provided with the product and Warning Label(s) on the product, we will repair or replace at our option, free of charge, upon its prepaid return to Mitutoyo with this WARRANTY card. ※ Purchaser is requested to fill the blank above and retain the warranty card at hand. Return card only when requesting Warranty Service.  Quality Assurance Department T. Narumi Manager This warranty shall not apply if the product and consumables have been subject to fair wear and tear, abuse through misuse or improper use/handling/storage/maintenance/service/repair or through adaptation/modification by the original purchase or any third party without prior written consent of Mitutoyo or as a result of damage by an actual disaster or circumstances beyond the control of Mitutoyo. To obtain service under this warranty the product must be returned to the store/dealer you purchased from along with the warranty card. Any postage, insurance, or shipping charges incurred in returning the product for service are the responsibility of the purchaser. ※ This warranty is not transferable and is only valid within the country of the original purchase. ※ You may have additional rights under the laws of country of original purchase that do not allow the exclusion of implied warranties of the exclusion or limitation of certain damages. If these laws apply, Mitutoyo's limitations and exclusions may not apply to you.			

SERVICE NETWORK

*As of February 2025

Europe

Mitutoyo Europe GmbH

Borsigstrasse 8-10, 41469 Neuss, GERMANY
TEL: 49 (0)2137 102-0 FAX: 49 (0)2137 102-351

Mitutoyo CTL Germany GmbH

Von-Gunzert-Strasse 17, 78727 Oberndorf, GERMANY
TEL: 49 (0)7423 8776-0 FAX: 49 (0)7423 8776-99

KOMEG Industrielle Messtechnik GmbH

Zum Wasserwerk 3, 66333 Völklingen, GERMANY
TEL: 49 (0)6898 91110 FAX: 49 (0)6898 911100

Germany

Mitutoyo Deutschland GmbH

Borsigstrasse 8-10, 41469 Neuss, GERMANY
TEL: 49 (0)2137 102-0 FAX: 49 (0)2137 86 85

M³ Solution Center Hamburg

Tempowerkring 9-im HIT-Technologiepark 21079
Hamburg, GERMANY
TEL: 49 (0)40 791894-0 FAX: 49 (0)40 791894-50

M³ Solution Center Berlin

Ernst-Lau-Straße 6, 12489 Berlin, GERMANY
TEL: 49 (0)30 2611 267 FAX: 49 30 67988729

M³ Solution Center Eisenach

Neue Wiese 4, 99817 Eisenach, GERMANY
TEL: 49 (0)3691 88909-0 FAX: 49 (0)3691 88909-9

M³ Solution Center Ingolstadt

Marie-Curie-Strasse 1A, 85055 Ingolstadt, GERMANY
TEL: 49 (0)841 954920 FAX: 49 (0)841 9549250

M³ Solution Center Leonberg

Am Längenbühl 3, 71229 Leonberg, GERMANY
TEL: 49 (0)7152 6080-0 FAX: 49 (0)7152 608060

Mitutoyo Deutschland GmbH - Small Tool Sales Division

Am Längenbühl 4, 71229 Leonberg, GERMANY
TEL: 49 (0)7152 9237-0 FAX: 49 (0)7152 9237-29

U.K.

Mitutoyo (UK) Ltd. HQ

Joule Road, West Point Business Park,
Andover, Hampshire SP10 3UX, UNITED KINGDOM
TEL: 44 (0)1264 353123 FAX: 44 (0)1264 354883

Coventry M³ Solution Centre

Unit6, Banner Park, Wickmans Drive, Coventry,
West Midlands CV4 9XA, UNITED KINGDOM
TEL: 44 (0)2476 426300

Halifax M³ Solution Centre

Lowfields Business Park, Navigation Close, Elland,
West Yorkshire HX5 9HB, UNITED KINGDOM
TEL: 44 (0)1422 375566

East Kilbride M³ Solution Centre

The Bairds Building, Rankine Avenue, Scottish
Enterprise Technology Park, East Kilbride G75
0QF, UNITED KINGDOM
TEL: 44 (0)1355 581170

France

Mitutoyo France

Paris Nord 2-123 rue de la Belle Etoile, BP 59267
ROISSY EN FRANCE 95957 ROISSY CDG
CEDEX, FRANCE
TEL: 33 (0)149 38 35 00

M³ Solution Center LYON

Parc Mail 523, cours du 3ème millénaire, 69791
Saint-Priest, FRANCE
TEL: 33 (0)149 38 35 70

M³ Solution Center STRASBOURG

Parc de la porte Sud, Rue du pont du péage,
67118 Geispolsheim, FRANCE
TEL: 33 (0)149 38 35 80

M³ Solution Center CLUSES

290 Avenue des Lacs, 74950 Scionzier, FRANCE
TEL: 33 (0)1 49 38 35 90

M³ Solution Center TOULOUSE

Aeroparc Saint Martin Cellule B08 ZAC de Saint
Martin du Touch 12 rue de Caulet 31300
Toulouse, FRANCE
TEL: 33 (0)1 49 38 42 90

M³ Solution Center RENNES

2, rue Claude Chappe, PA le Vallon - ZAC Mivoie,
35230 Noyal-Châtillon-sur-Seiche, FRANCE
TEL: 33 (0)1 49 38 42 10

Italy

Mitutoyo Italiana S.r.l.

Corso Europa, 7 - 20045 Lainate (MI), ITALY
TEL: 39 02 935781 FAX: 39 02 93578255

M³ Solution Center BOLOGNA

Via dei Carpini 1/A - 40011 Anzola Emilia (BO), ITALY
TEL: 39 02 93578215 FAX: 39 02 93578255

M³ Solution Center CHIETI

Contrada Santa Calcagna - 66020 Rocca S.
Giovanni (CH), ITALY
TEL: 39 02 93578280 FAX: 39 02 93578255

M³ Solution Center PADOVA

Via G. Galilei 21/F - 35035 Mestrino (PD), ITALY
TEL: 39 02 93578268 FAX: 39 02 93578255

Netherlands**Mitutoyo Nederland B.V.**

Storkstraat 30, 3905 KX Veenendaal,
THE NETHERLANDS
TEL: 31(0)318-534911

Mitutoyo Nederland B.V. / M³ Solution Center Enschede

Institutenweg 50, 7521 PK Enschede,
THE NETHERLANDS
TEL: 31(0)318-534911

Mitutoyo Nederland B.V. / M³ Solution Center Eindhoven

De Run 1115, 5503 LB Veldhoven,
THE NETHERLANDS
TEL: 31(0)318-534911

Mitutoyo Research Center Europe B.V.

De Rijn 18, 5684 PJ Best, THE NETHERLANDS
TEL: 31(0)499-320200 FAX: 31(0)499-320299

Belgium**Mitutoyo Belgium N.V. / M³ Solution Center Melsele**

Schaarbeekstraat 20, B-9120 Melsele, BELGIUM
TEL: 32 (0)3-2540444

Sweden**Mitutoyo Scandinavia AB**

Släntvägen 6, 194 61 Upplands Väsby, SWEDEN
TEL: 46 (0)8 594 109 50

Mitutoyo Scandinavia AB / M³ Solution Center Alingsås

Ängsvaktaregatan 3A, 441 38 Alingsås, SWEDEN
TEL: 46 (0)8 594 109 50

Mitutoyo Scandinavia AB / M³ Solution Center Värnamo

Kalkstensvägen 7, 331 44 Värnamo, SWEDEN
TEL: 46 (0)8 594 109 50

Switzerland**Mitutoyo (Schweiz) AG**

Steinackerstrasse 35, 8902 Urdorf, SWITZERLAND
TEL: 41 (0)447361150

Mitutoyo (Suisse) SA

Rue Gallée 4, 1400 Yverdon-les Bains, SWITZERLAND
TEL: 41 (0)244259422

Poland**Mitutoyo Polska Sp.z o.o.**

Ul.Graniczna 8A, 54-610 Wrocław, POLAND
TEL: 48 (0)71354 83 50 FAX: 48 (0)71354 83 55

Czech Republic**Mitutoyo Česko s.r.o.**

Dubská 1626, 415 01 Teplice, CZECH REPUBLIC
TEL: 420 417-514-011 Email: info@mitutoyo.cz

Mitutoyo Česko s.r.o. M³ Solution Center Ivančice

Ke Karlovu 62/10, 664 91 Ivančice, CZECH REPUBLIC
TEL: 420 417-514-011 Email: info@mitutoyo.cz

Mitutoyo Česko s.r.o. M³ Solution Center Ostrava Mošnov

Mošnov 314, 742 51 Mošnov, CZECH REPUBLIC
TEL: 420 417-514-050 Email: info@mitutoyo.cz

Mitutoyo Česko s.r.o. Slovakia Branch

Šoltésvej 346/1, 017 01 Považská Bystrica,
SLOVAKIA
TEL: 421 948-595-590 Email: info@mitutoyo.sk

Hungary**Mitutoyo Hungária Kft.**

Galamb József utca 9, 2000 Szentendre, HUNGARY
TEL: 36 (30) 6410210

Romania**Mitutoyo Romania SRL**

1A Drumul Garii Odai Street, showroom, Ground
Floor, 075100 OTOPENI-ILFOV, ROMANIA
TEL: 40 (0)311012088 FAX: +40 (0)311012089

Showroom in Brasov

Strada Ionescu Crum Nr.1, Brasov Business Park
Turnul 1, Mezanin, 500446
Brasov-Judetul Brasov, ROMANIA
TEL/FAX: 40 (0)371020017

Finland**Mitutoyo Scandinavia AB Finnish Branch**

Autokeskuksentie 16B, 33960 Pirkkala, FINLAND
TEL: 358 (0)40 355 8498

Austria**Mitutoyo Austria GmbH**

Salzburger Straße 260/2 und 260/3, 4600 Wels,
AUSTRIA
TEL: 43 (0)7242 219 998

Mitutoyo Austria GmbH Goetzis Regional showroom

Lastenstrasse 48a, 6840 Götzis, AUSTRIA

Singapore

Mitutoyo Asia Pacific Pte. Ltd.

Head office / M³ Solution Center

24 Kallang Avenue, Mitutoyo Building,
SINGAPORE 339415

TEL:(65)62942211 FAX:(65)62996666

Malaysia

Mitutoyo (Malaysia) Sdn. Bhd.

Kuala Lumpur Head Office / M³ Solution Center

Mah Sing Integrated Industrial Park, 4, Jalan Utarid U5/14,
Section U5, 40150 Shah Alam, Selangor, MALAYSIA

TEL:(60)3-78459318 FAX:(60)3-78459346

Penang Branch office / M³ Solution Center

30, Persiaran Mahsuri 1/2, Sunway Tunas, 11900
Bayan Lepas, Penang, MALAYSIA

TEL:(60)4-6411998 FAX:(60)4-6412998

Johor Branch office / M³ Solution Center

70 (Ground Floor), Jalan Molek 1/28, Taman
Molek, 81100 Johor Bahru, Johor, MALAYSIA

TEL:(60)7-3521626 FAX:(60)7-3521628

Thailand

Mitutoyo (Thailand) Co., Ltd.

Bangkok Head Office / M³ Solution Center

2327 Onnut Road Kwaeng Onnut Khet Suanluang
Bangkok 10250, THAILAND

TEL:(66)2080 3500

Chonburi Branch / M³ Solution Center

7/1, Moo 3, Tambon Bowin, Amphur Sriracha,
Chonburi 20230, THAILAND

TEL:(66)2080 3563 FAX:(66)3834 5788

ACC Branch / M³ Solution Center

122/8, 122/9, Moo 6, Tambon Donhuaroh, Amphur
Muangchonburi, Chonburi 20000, THAILAND

TEL:(66)2080 3565

Indonesia

PT. Mitutoyo Indonesia

Head Office / M³ Solution Center

Jalan Sriwijaya No.26 Desa cibatu Kec. Cikarang
Selatan Kab. Bekasi 17530, INDONESIA

TEL: (62)21-2962 8600 FAX: (62)21-2962 8604

Batam Branch Office

Business Center Adhya Building 3rd Floor Kom-
pleks Permata Niaga Blok A No. 1, Jalan jendral
Sudirman Kelurahan Sukajadi, Kecamatan Batam
Kota, Kepulauan Riau 29444, INDONESIA

TEL: (62)-778-4888000

Vietnam

Mitutoyo Vietnam Co., Ltd

Hanoi Head Office / M³ Solution Center

1st & 2nd floor, MHDI Building, No. 60 Hoang Quoc
Viet Road, Nghia Do Ward, Cau Giay District, Hanoi,
VIETNAM

TEL:(84)24-3768-8963 FAX:(84)24-3768-8960

Ho Chi Minh City Branch Office / M³ Solution Center

Unit No. B-00.07, Ground Floor, C1 Building, No.
6, Street D9, An Loi Dong Ward, Thu Duc City, Ho
Chi Minh City, VIETNAM

TEL:(84)28-3840-3489 FAX:(84)28-3840-3498

Hai Phong City Branch Office

Room 511, 5th Floor, Thanh Dat 3 Building, No. 4
Le Thanh Tong Street, May To Ward, Ngo Quyen
District, Hai Phong City, VIETNAM

TEL:(84)22-5398-9909

Philippines

Mitutoyo Philippines, Inc.

Head Office / M³ Solution Center

Unit 1B & 2B LTI, Administration Building 1, Annex 1, North
Main Avenue, Laguna Technopark, Binan Laguna 4024,
PHILIPPINES

TEL/FAX:(63) 49 544 0272

India

Mitutoyo South Asia Pvt. Ltd. Head Office

C-122, Okhla Industrial Area, Phase-I,
New Delhi-110 020, INDIA

TEL: (91) 11-40578485/86

MSA Technical Center

Plot no. 65, Ground Floor, Udyog Vihar, Phase-4 Gurga-
on, Haryana - 122016, INDIA

TEL : (91) 124-2340286/287

Mumbai Region Head office

303, Sentinel Hiranandani Business Park Powai,
Mumbai-400 076, INDIA

TEL: (91) 22-25700684/685/837/839

Pune Office / M³ Solution Center

G4/G5, Pride Kumar Senate, Off. Senapati Bapat
Road, Pune-411 016, INDIA

TEL:(91) 20-25660043/44/45

Ahmedabad Office / M³ Solution Center

A-104 & A-105, First Floor, Solitaire Corporate
Park, Near Divya Bhaskar Press, S.G. Road,
Ahmedabad - 380 015, INDIA

TEL: (91) 079 - 29704902/903

Bengaluru Region Head office / M³ Solution Center

116/117-2, Ground Floor, Sy. No. 93 & 94, 3rd
Phase, Peenya Industrial Area, Bengaluru-560
058, INDIA

TEL: (91) 80-25630946/47/48/49

Coimbatore Office

Regus, Srivari Srimath, 3rd Floor, Door No:1045,
Avinashi Road, Coimbatore - 641 018,INDIA
TEL: (91) 9345005663

Chennai Office / M³ Solution Center

No. 624, Anna Salai Teynampet, Chennai-600 018, INDIA
TEL: (91) 44-24328823/24/25

Kolkata Office

Unit No. 1208,Om Tower, 32,J.L.Nehru Road,
Kolkata-700 071, INDIA
TEL: (91) 33-22267088/40060635/22266817

Taiwan

Mitutoyo Taiwan Co., Ltd. / M³ Solution Center Taipei
4F., No.71, Zhouzi St., Neihs Dist.,Taipei City 114,
TAIWAN
TEL:886(2)5573-5900 FAX:886(2)8752-3267

Taichung Branch / M³ Solution Center Taichung
1F., No. 299, Gaotie 1st Rd., Wuri Dist., Taichung
City 414, TAIWAN
TEL:886(4)2338-6822 FAX:886(4)2338-6722

Kaohsiung Branch / M³ Solution Center Kaohsiung
1F., No.31-1, Haibian Rd., Lingya Dist.,
Kaohsiung City 802, TAIWAN
TEL:886(7)334-6168 FAX:886(7)334-6160

South Korea

Mitutoyo Korea Corporation

Head Office / M³ Solution Center

33, Eungyejungang-ro 306beon-gil, Siheung-si,
Gyeonggi-do, 15120 KOREA
TEL:82(31)361-4200 FAX:82(31)361-4201

Busan Office / M³ Solution Center

(3150-3, Daejeo 2-dong) 8,Yutongdanji 1-ro
49beon-gil, Gangseo-gu, Busan, 46721 KOREA
TEL:82(51)324-0103 FAX:82(51)324-0104

Daegu Office / M³ Solution Center

(Galsan-dong, Daegu Business Center), 301-Ho, 217,
Seongseogongdan-ro, Dalseo-gu, Daegu 42704 KOREA
TEL:82(53)593-5602 FAX:82(53)593-5603

China

Mitutoyo Measuring Instruments (Shanghai) Co., Ltd.

18/F, NEW BUND Shun Tak Center, No.18, Lane
666, West Haiyang Road, Pudong New District,
Shanghai 200124, CHINA
TEL:86(21)5836-0718 FAX:86(21)5836-0717

Suzhou Office / M³ Solution Center China (Suzhou)

1/2 Floor, Building 4, No.175 Songbei Road,
Suzhou Free Trade Zone, Suzhou City, Jiangsu
215000, CHINA
TEL:86(512)6522-1790 FAX:86(512)6251-3420

Wuhan Office / M³ Solution Corner

Room 1701, Wuhan Wanda Center, No. 96, Linji-
ang Road, Wuchang District, Wuhan
Hubei 430060, CHINA
TEL:86(27)8544-8631 FAX:86(27)8544-6227

Chengdu Office

Room 1-102, 1st Floor, Unit 1, Building 1, No. 24,
Wannian Road (Wanniancang Cool), Chenghua
District, Chengdu City, Sichuan 610056, CHINA
TEL:86(28)8671-8936 FAX:86(28)8671-9086

Hangzhou Office

Room 329, 3F, Building D, West Square of Hang-
zhoudong Railway Station, No. 260 Xinfeng Road,
Shangcheng District, Hangzhou City, Zhejiang
310002, CHINA
TEL: 86(571)8288-0319 FAX: 86(571)8288-0320

Tianjin Office / M³ Solution Center China (Tianjin)

Room D 12/F, TEDA Building, No.256 Jie-fang
Nan Road Hexi District,Tianjin 300042, CHINA
TEL:86(22)5888-1700 FAX:86(22)5888-1701

Changchun Office

Room 815, 8F, Building A1, Upper East
International No.3000 Dongsheng Street,
Erdao District, Changchun, Jilin, 130031, CHINA
TEL:86(431)8192-6998 FAX:86(431)8192-6998

Chongqing Office

Room 1312, Building 3, Zhongyu Plaza, No.86,
Hongjin Avenue,Longxi Street, Yubei District,
Chongqing, 400000, CHINA
TEL:86(23)6595-9950 FAX:86(23)6595-9950

Qingdao Office

Room 638, 6F, No.192 Zhengyang Road, Chengyang
District, Qingdao, Shandong, 266109, CHINA
TEL:86(532)8096-1936 FAX:86(532)8096-1937

Xi'an Office

Room 1503 Jianke Building, No.99 Yanta Road,
Beilin District, Xi'an City, Shaanxi 710054, CHINA
TEL:86(29)8538-1380 FAX:86(29)8538-1381

Dalian Office / M³ Solution Center China (Da- lian)

Room A-106 Shuijing SOHO, No.16 Harbin Road,Eco-
nomic Development Zone,Dalian, 116600 CHINA
TEL:86(411)8718 1212 FAX:86(411)8754-7587

Zhengzhou Office

Room 5003, 50th Floor, South Tower of Greenland
Centre, No.36 Yulin North Road, Zhengdong New
District, Zhengzhou, Henan, 450000, CHINA
TEL:86(371)6097-6436 FAX:86(371)6097-6981

Dongguan Office / M³ Solution Center China (Dongguan)

Room 801, No 65, Chang'an Section Guanchang Road, Chang'an Town, Dongguan City, Guangdong 523841, CHINA
TEL:86(769)8541 7715 FAX:86(769)-8541 7745

Fuzhou Office

Unit 03, 7th floor of East Tower, Sansheng International Center, No.118 Wusi Road, Gulou District, Fuzhou City, Fujian 350001, CHINA
TEL: 86 (591) 8761 8095
FAX: 86 (591) 8761 8096

Changsha Office

Room 2207, Building 1, Shiner International Plaza, No. 88, Kaiyuan Middle Road, Changsha City, Hunan 410100, CHINA
TEL: 86 (731) 8401 9276
FAX: 86 (731) 8401 9376

Changzhou Office

Room 1502, Joint Financial Tower, No.255, Tongjiang North Road, Tianning District, Changzhou City, Jiangsu 2130002, CHINA
TEL:86(519)8815 8319 FAX:86(519)8815 8319

Wenzhou Office

Room 512, Building 4, Xinjingdujiayuan, Sanyang Street, Ouhai District, Wenzhou City, Zhejiang 325014, CHINA
TEL:86(577)8641 5280

Shunde Office

Room 1603, Building 26, Vanke Golden Riverside Plaza Phase II, No.13 Mid DeSheng Road, Shunde District, Foshan City, Guangdong 528300, CHINA
TEL/FAX: 86(757)2228 8621

Hefei Office

UnitB3111-1, 1F, Block B3, Huiyuan, Heping Road, Yaohai District, Hefei City, Anhui 230001, CHINA
TEL: 86(551)6560 1006

Mitutoyo Measuring Technology (Suzhou) Co., Ltd.

1/2 Floor, Building 4, No.175 Songbei Road, Suzhou Free Trade Zone, Suzhou City, Jiangsu 215000, CHINA
TEL:86(512)6252-2660 FAX:86(512)6252-2580

USA**Mitutoyo America Corporation**

965 Corporate Blvd., Aurora, IL 60502, U.S.A.
TEL:1-(630)820-9666 Toll Free No. 1-888-648-8869
FAX:1-(630)978-3501

Headquarters (Aurora) / M³ Solution Center

965 Corporate Blvd., Aurora, IL 60502, U.S.A.

Seattle (Renton) Office / M³ Solution Center

1000 SW 34th St. Suite G, Renton, WA 98057 U.S.A.
TEL:1-(888)-648-8869

Houston Office / M³ Solution Center

4560 Kendrick Plaza Drive Suite 120 Houston, TX 77032, U.S.A.
TEL:1-(888)-648-8869 FAX:1-(281)227-0937

Cincinnati (Mason) Office / M³ Solution Center

6220 Hi-Tek Ct., Mason, OH 45040, U.S.A.
TEL:1-(888)-648-8869 FAX:1-(513)754-0718

Detroit (Novi) Office / M³ Solution Center

46850 Magellan Drive, Suite 100 Novi, MI 48377, U.S.A.
TEL:1-(888)-648-8869 FAX: 1-(248)-926-0928

Los Angeles (City of Industry) Office / M³ Solution Center

16925 E. Gale Ave., City of Industry, CA 91745, U.S.A.
TEL:1-(888)-648-8869 FAX:1-(626)369-3352

Charlotte (Huntersville) Office / M³ Solution Center

11515 Vanstory Dr., Suite 140, Huntersville, NC 28078, U.S.A.
TEL:1-(888)-648-8869 FAX:1-(704)875-9273

Boston (Marlborough) Office / M³ Solution Center

753 Forest Street, Suite 110, Marlborough, MA 01752, U.S.A.
TEL:1-(888)648-8869 FAX:1-(508)485-0782

West Chester (Ohio) / Metrology Service Center

8876 Beckett road, West Chester, OH 45069, USA
TEL:1-(888)-648-8869

Mitutoyo America Corporation Calibration Lab

965 Corporate Blvd., Aurora, IL 60502, U.S.A.
TEL:1-(888)-648-8869 FAX:1-(630)978-6477

Mitutoyo America Corporation CT-Lab Chicago

965 Corporate Blvd., Aurora, IL 60502, U.S.A.
TEL: 1-(888)-648-8869 FAX: 1-(630)-820-3418

Mitutoyo Research & Development America, Inc.

11533 NE 118th St., Kirkland,
WA 98034-7111, U.S.A.
TEL:1-(425)821-3906 FAX:1-(425)821-32280

Mitutoyo Research & Development America, Inc. - California Office

16925 Gale Ave. City of Industry,
CA 91745-1806 U.S.A.
TEL: 1-(425)821-3906 FAX: 1-(425)821-3228

Canada

Mitutoyo Canada Inc.

2121 Meadowvale Blvd., Mississauga,
Ont. L5N 5N1., CANADA
TEL:1-(905)821-1261 FAX:1-(905)821-4968

Montreal Office

7075 Place Robert-Joncas Suite 129, Montreal,
Quebec H4M 2Z2, CANADA
TEL:1-(514)337-5994 FAX:1-(514)337-4498

Brazil

Mitutoyo Sul Americana Ltda.

Head office / M³ Solution Center

Avenida Mimes nº 25 – Loteamento Multivias II,
Jardim Ermida I, CEP 13212-216 Jundiaí - SP,
BRASIL
TEL: 55 (11) 5643-0004/0041

Filial Curitiba / M³ Solution Center

Rua Sergipe, nº 101, Sala A, Bairro Boneca do
Iguaçu, São José dos Pinhais – Paraná – BRA-
SIL CEP 83040120
TEL: 55 (41) 3534-1728

Filial Joinville / M³ Solution Center

Rua Sorocaba, No.265, Bairro Floresta, Join-
ville, Santa Catarina, BRASIL, CEP 89212-210
TEL: 55 (47) 3025-8062

Filial Porto Alegre / M³ Solution Center

Rua Emilio Boeckel nº 325, Bairro: Flao, São
Leopoldo - RS, BRASIL CEP 93020-600
TEL: 55 (51) 3037-9383

Argentina

Mitutoyo Sul Americana Ltda.

Argentina Branch / M³ Solution Center

AV. MITRE 1249 Piso. PB– FLORIDA – BUENOS
AIRES ARGENTINA
TEL:54 (11) 4730-1433 FAX:54 (11) 4730-1411

Mexico

Mitutoyo Mexicana, S.A. de C.V.

Industria Eléctrica No.15, Parque Industrial, Nau-
calpan de Juárez, Estado de México C.P.53370,
MÉXICO

TEL: 52 (01-55) 5312-5612
FAX: 52 (01-55) 5312-3380

Monterrey Office / M³ Solution Center

Blv. Interamericana No. 103, Parque Industrial
FINSA, C.P. 66636 Apodaca, N.L., MÉXICO
TEL: 52(01-81) 8398-8227/8228/8242/8244
FAX: 52(01-81) 8398-8226

Tijuana Office / M³ Solution Center

Calle José María Velazco 10501-C, Col. Cd. Industrial
Nueva Tijuana, C.P. 22500 Tijuana, B.C., MÉXICO
TEL: 52 (01-664) 647-5024

Querétaro Office / M³ Solution Center

Av. Cerro Blanco No.500-1, Colonia Centro Sur,
Querétaro, Querétaro, C.P. 76090, MÉXICO
TEL: 52 (01-442) 340-8018, 340-8019 and 340-8020
FAX: 52 (01-442) 340-8017

**Mitutoyo Mexicana, S.A. de C.V. Querétaro
Calibration Laboratory**

Av. Cerro Blanco 500 30 Centro Sur,
Querétaro, Querétaro, C.P. 76090, MÉXICO
TEL: 52 (01-442) 340-8018, 340-8019 and 340-8020
FAX: 52 (01-442) 340-8017

Aguascalientes Office / M³ Solution Center

Av. Aguascalientes No. 622, Local 15 Centro Comer-
cial El Cilindro Fracc. Pulgas Pandas Norte, C.P.
20138, Aguascalientes, Ags. MÉXICO
TEL: 52 (01-449) 174-4140 and 174-4143

Irapuato Office / M³ Solution Center

Boulevard a Villas de Irapuato No. 1460 L.1 Col. Ejido
Irapuato C.P. 36643
Irapuato, Gto., MÉXICO
TEL: 52 (01-462) 144-1200 and 144-1400

Revision Record

Date of publication	Revision status	Details of revision
June 1, 2025	First edition	Publication

Mitutoyo Corporation

20-1, Sakado 1-Chome, Takatsu-ku, Kawasaki-shi, Kanagawa 213-8533, Japan
Tel: +81 (0)44 813-8230 Fax: +81 (0)44 813-8231
Home page: <https://www.mitutoyo.co.jp/global.html>

For the EU Directive, Authorized representative and importer in the EU:
Mitutoyo Europe GmbH
Borsigstrasse 8-10, 41469 Neuss, Germany

For the UK Regulation, Authorized representative and importer in the UK:
Mitutoyo (UK) Ltd.
Joule Road, West Point Business Park, Andover, Hampshire SP10 3UX, UNITED KINGDOM